


PEF1/E



P r o c e d u r e :

For a proper use ensure a constant flow into the probe holder PEF1/E.

- Connect water input and output to the probe holder.
- Regulate the flow, by turning the regulation knob, so that the floater will reach the level line.
- Connect the SEPR to the instrument. The SEPR will send a signal to the instrument.
- The instrument will start to receive probe's data.
- Note: The instrument doesn't receive any information from the probe if there isn't flow into the probe holder.